

PCB Piezotronics Inc. claims proprietary rights in the information disclosed hereon. Neither it nor any reproduction thereof will be disclosed to others without the written consent of PCB Piezotronics Inc.


REVISIONS

REV	DESCRIPTION	ECN	APP'D
NR	RELEASED TO DRAFTING		


38072

Related Drawing

No modifications permitted without the approval of the authorized person



MOUNTING HOLE PREPERATION
Ø.332[8.43]
THRU
1/8-27 NPT



- 2 ABOVE INSTALLATION IS FOR WALL THICKNESS OF .33[8.4]. L₁Ø.750[19.05] FOR THICKER WALLS TO CLEAR 7/16[11.11] SOCKET.
- 1 RECOMMENDED MOUNTING TORQUE ON A 7/16[11.11] HEX: 5 FOOT POUNDS[7 NEWTON METERS]

UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:		DRAWN	FCB	7/25/07	MFG	P.R.R	7-25-07	<p>3425 WALDEN AVE. DEPEW, NY 14043 (716) 684-0001 E-MAIL: sales@pcb.com</p>
DIMENSIONS IN INCHES	DIMENSIONS IN MILLIMETERS [IN BRACKETS]	CHK'D	ECB	7/25/07	ENGR	RF	7-25-07	
DECIMALS XX ± .01 XXX ± .005	DECIMALS X ± 0.3 XX ± 0.13	APP'D	NJL	7/25/07	SALES	RLM	7-25-07	
ANGLES ± 2 DEGREES	ANGLES ± 2 DEGREES	TITLE						
FILLETS AND RADII .003 - .005	FILLETS AND RADII [0.07 - 0.13]	INSTALLATION DRAWING MODELS 102A43,44,45 PRESSURE SENSOR						CODE IDENT. NO. 52681
								DWG. NO. 38072
								SCALE: 2X
								SHEET 1 OF 1