ľ	Model	Number
	377	7C 10

PRECISION CONDENSER MICROPHONE

Revision: K ECN #: 50697

ENGLISH	SI	
1/4"	1/4"	
Pressure	Pressure	
- 60 dB re 1 V/Pa	- 60 dB re 1 V/Pa	[1]
1.0 mV/Pa	1.0 mV/Pa	[1]
4 to 70,000 Hz	4 to 70,000 Hz	
0.25 to 3.0 Hz	0.25 to 3.0 Hz	
65,000 Hz	65,000 Hz	[2]
< 35 dB(A) re 20 µPa	< 35 dB(A) re 20 μPa	
174 dB re 20 μPa	174 dB re 20 μPa	[2]
-40 to +248 °F	-40 to +120 °C	
0.004 dB/°F	0.007 dB/°C	[2][1]
-0.003 dB/kPa	-0.003 dB/kPa	[2][1]
± 0.0012 dB/%RH	± 0.0012 dB/%RH	[1]
66 dB re 20 μPa	66 dB re 20 μPa	[2]
5 pF	5 pF	[2]
0 V	0 V	[3]
Stainless Alloy	Stainless Alloy	
Rear	Rear	
0.2244 - 60 UNS	5.7 mm - 60 UNS	
0.25 - 60 UNS	6.35 mm - 60 UNS	
0.28 in x 0.41 in	7.0 mm x 10.5 mm	
0.25 in x 0.36 in	6.4 mm x 9.1 mm	
0.06 oz	1.8 gm	[2]
	-	
	1/4" Pressure - 60 dB re 1 V/Pa 1.0 mV/Pa 4 to 70,000 Hz 0.25 to 3.0 Hz 65,000 Hz < 35 dB(A) re 20 μPa 174 dB re 20 μPa -40 to +248 °F 0.004 dB/°F -0.003 dB/kPa ± 0.0012 dB/%RH 66 dB re 20 μPa 5 pF 0 V Stainless Alloy Rear 0.2244 - 60 UNS 0.25 - 60 UNS 0.28 in x 0.41 in 0.25 in x 0.36 in	1/4"

OPTIONAL VERSIONS

Optional versions have identical specifications and accessories as listed for the standard model except where noted below. More than one option may be used.

NOTES:

[1]re 250 Hz

[2]Typical. [3]Prepolarized

[4] See PCB Declaration of Conformance PS167 for details.

SUPPLIED ACCESSORIES:

Model ACS-20 Calibration of Precision Condensor Microphones (1)

Entered: LK Engineer: SPW Sales: MV Approved: AJA Spec Number: Date: 05/27/2020 Date: 05/27/2020 Date: 05/27/2020

> Phone: 716-684-0001
> Fax: 716-684-0987
> E-Mail: info@pcb.com 3425 Walden Avenue, Depew, NY 14043

Date: 05/27/2020

40215

All specifications are at room temperature unless otherwise specified. In the interest of constant product improvement, we reserve the right to change specifications without notice. ICP® is a registered trademark of PCB Piezotronics, Inc.